



Kjærulf
Pedersen a/s

Consignee

PR Electronics A/S
Lerbakken 10
8410 Rønne
Denmark

Consignor

Kjaerulf Pedersen a/s
Taastrupgaardsvej 8-10
2630 Taastrup
Denmark

Taastrup, 22/4 2015

Certificate of Conformity

KP order No.....: 146208
Customer order No.....: 1179624
KP item No.....: KP5456135-3
Customer item No.....: ---
Resistance thermometer type.....: RT type AUI, 1xPt100 cl. B 1/6 DIN, 3-wire,
W.No. 1.4571 Ø9x1x300 mm,
B-head Ø15.3 mm w/cable entry M20x1.5,
prepared for transmitter, max. 250°C
Resistor 1/6 DIN EN60751 cl. B.....: RoHS compliant
Resistivity: 100 Ohm at 0°C
Tolerance: Offset: 1/6 DIN EN 60751 class B
TCR: DIN EN 60751 class B
T.-Range: -200°C - +600°C
Dimension: 2.3 x 2 x 1.3 mm (LxWxH)
Contacts: Pt/Ni-wire, Ø0.2 mm, 10 mm long
The 1/6 (or +/- 50mK @ 0°C plus 0.8mK/K)
is valid between ~-10°C up to ~+100°C
Produced by.....: Kjaerulf Pedersen a/s
Taastrupgaardsvej 8-10
2630 Taastrup
Denmark
Conform to the following standard(s).....: Pt100 resistance thermometer acc. to IEC 60751

We certify that the above mentioned items correspond to your order and to what were agreed upon in our order confirmation of same.

Kind regards
Kjaerulf Pedersen a/s

Kjærulf Pedersen a/s
Taastrupgaardsvej 8-10
2630 Taastrup

Michael Steenberg

